

Certificate of Quality

Product Description: TipOne® Pipette Tips and TipOne® Filter Tips

Tip Material: 100% virgin polypropylene

USA Scientific certifies that this product is produced in a fully controlled environment that is designed for the product to be free from detectable RNase, DNase, DNA and Endotoxins. Selected samples of product from this production environment have been tested for the following parameters on a recurring basis.

	DNasa	
•	RNase	
	Testing Method:	RNA Digestion
	Limit:	< 1 x 10 ^{.9} Kunitz Units
•	DNase	
	Testing Method:	DNA Digestion
	Limit:	< 1 x 10 ⁻ Kunitz Units
•	Human DNA	
	Testing Method:	Polymerase Chain Reaction
	Limit:	< 2 x 10 ^{.9} mg; less than one human cell
•	Endotoxin	
	Testing Method:	Kinetic-chromogenic LAL-test (Ph. EUR. 2.6.1
	Limit:	<0.05 EU/ml
	Limit:	<0.05 EU/ml

The concentration levels of the tested samples for RNase, DNAse, DNA and Endotoxins are less than the above stated limits.

4)